



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Superior Scale, Inc.
2118 Carolina Place
Fort Mill, SC 29708

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 19 March 2023
Certificate Number: L2077-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Superior Scale, Inc.

2118 Carolina Place
Fort Mill, SC 29708
Steve Daniels
803-548-3320

CALIBRATION

Valid to: **March 19, 2023**

Certificate Number: **L2077-1**

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Thermocouple Simulation Type RTD Type J Type K Type T	(-320 to 1 500) °F (-320 to 1 350) °F (-300 to 2 500) °F (-300 to 750) °F	0.56 °F 1 °F 1.1 °F 1.3 °F	Process Calibrator w/ Thermocouple Probe

Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method and/or Equipment
Pin/Plug Gages	(0 to 1) in	(106 + 25L) μin	Bench Micrometer
Calipers	(0 to 60) in	(569 + 32L) μin	Gage Blocks
Micrometers	(0 to 4) in (0 to 24) in (0 to 36) in	(62 + 43L) μin (139 + 63L) μin (961 + 18L) μin	
Indicators	(0 to 0.5) in (0 to 6) in	663 μin (613 + 81L) μin	
Heights Gages	(0 to 48) in	(595+15L) μin	Gage Blocks and Surface Plate
Rulers	(0 to 36) in	0.015 in	Gage Blocks and Lenscope

Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Analytical Balances (0.000 1 g Resolution)	(0 to 200) g	0.85 mg	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(0.001 g Resolution)	(0 to 380) g	2.15 mg	
(0.01 g Resolution)	(0 to 1 000) g	17.92 mg	
Lab Balances (0.1 g Resolution)	(0 to 4.1) kg	0.15 g	ASTM E617 Class 2 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
Bench Scales (0.002 lb Resolution)	(0 to 60) lb	0.009 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(0.01 lb Resolution)	(0 to 100) lb	0.019 lb	
(0.05 lb Resolution)	(0 to 500) lb	0.094 lb	
Truck Scales (20 lb Resolution)	(0 to 300 000) lb	30 lb	
Industrial Scales ³ (0.1 lb Resolution)	(0 to 1 000) lb	0.19 lb	
(0.2 lb Resolution)	(0 to 2 000) lb	0.37 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(0.5 lb Resolution)	(0 to 5 000) lb	0.93 lb	
(1 lb Resolution)	(0 to 10 000) lb	1.9 lb	
(2 lb Resolution)	(0 to 20 000) lb	3.7 lb	

Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Industrial Scales ³ (5 lb Resolution)	(0 to 50 000) lb	9.3 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(10 lb Resolution)	(0 to 100 000) lb	19 lb	
Pressure Gages – Hydraulic	(0 to 300) psig	0.43 psig	Druck Model DPI611 Pressure Calibrator w/ 5000 PSIG External Transducer
	(0 to 5 000) psig	12 psig	
Pressure Gages – Vacuum	(-13 to 0) psiv	0.087 psiv + 0.2 % of reading	Druck Model DPI611

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Relative Humidity Indicators	(30 to 85) % RH	2.63% RH	Compared to Psychro-Dyne
Thermocouple and Resistance Temperature Probes and Systems	(0 to 300) °C	1.5 °C	Process Calibrator w/RTD Probe and Dry Block Calibrator

Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Timers	(0.5 to 60) min	0.77 sec	Reference Stopwatch
	(1 to 8) hours	1.63 sec	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = Length is in inches.
3. Industrial Scales include Floor, Tank, Hopper Crane, etc.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L2077-1.



R. Douglas Leonard Jr., VP, PILR SBU

